Art. 602.B200

Sabouraud Dextrose Agar + Chloramphenicol 200ml

	Technical Data Sheet
Use in	• Pharmaceutical Industry • Cosmetic Industry • For industrial, laboratory & research applications only
Use for	• Plate pour method • Identification and growth of yeasts and moulds
Typical composition per liter	Casein peptone 5g — Meat peptone 5g Glucose x H_2O* 44g — Chloramphenicol 50mg Agar 15g *Glucose-D(+)*H_2O = Glucose monohydrate *44g Glucose monohydrate = 40g Glucose = 40g Dextrose This medium can be adjusted / or supplemented according to the performance criteria required.
Filling volume	• 200ml
Bottle format	• 220ml screw cap bottle with 2 integrated septa • Colour of cap: red
Units per pack	• 12 bottles on a plastic tray wrapped with shrink foil
Shelf life	• 9 months from production date
Storage	• Recommended storage temperature: 2-25°C • Should be stored at temperatures as stable as possible
Label	• On the side • Contain autoclave indicator
Label information	• Product name: SDA + CA 200ml • Expiry date: YYYYMMMDD ➔ MMM in letters (e.g.: 2023Nov04) • Lot-number • Individual number • Barcode
Barcode	• 2-dimensional (data matrix), 20 digits: • Digits 1-3: Art.-No. • Digits 4-9: Lot-Number • Digits 10-14: Individual-Number • Digits 15-20: Date (YYMMDD)
Delivery	• Temperature controlled delivery on request • For shipments of larger amounts plastic pallets in Euro-size are used

Sabouraud Dextrose Agar + Chloramphenicol 200ml

	Technical Data Sheet
Bottle information	• Label contains autoclaving indicator (pink ➔ grey-brown) • Bottles are incubated for >48 hours at 30-35°C after production and then packed immediately • Bottles are not touched any more by hand after autoclaving
Melting instructions	• Open the lid slightly and melt the agar (e.g., at 90 to 95°C for about 45 min in a water bath or 105°C for 15 min in free-flowing steam) • Transfer the bottle to a water bath at ~47°C for a maximum time of 8h • Pipette your sample into an empty, sterile petri dish • Add 18-25 ml of agar and shake gently • Incubate the plates upside down • **Do not** use a microwave for melting the agar • **Do not** reuse melted agar a second time
Place of production	PharmaMedia Dr. Müller GmbH Gustav-Throm-Str. 1, 69181 Leimen - Germany

	Quality control, Certificates
Certificates	Each lot of product can be obtained with a certificate of analysis (CoA):

Physico-chemical test parameters:				
Appearance	Clear, yellowish			
pH value	5,4 – 5,8			
Filling volume	194 – 206ml			
Growth Promotion test: 10-100 CFU				
C.albicans	ATCC 10231	30-35°C	1 day	50-200%
C.albicans	ATCC 10231	20-25°C	2-3 d	50-200%
A.brasiliensis	ATCC 16404	20-25°C	3-5 d	50-200%
Inhibition test: 10.000-100.000 CFU				
E.coli	ATCC 8739	30-35°C	1 day	No growth
Sterility control				
≥ 7 days at 30-35°C, no growth				

Certificate of origin	All media lots produced by PMM can be obtained with a Certificate of Origin (CoO). All animal derived raw materials are specified as follows: • Raw material • Tissue • Animal source • Country of origin • Infectivity category (acc. to TSE guideline: EMA/410/01 rev. 3)

	Quality control, Certificates
BSE policy	• In compliance with the current note for guidance on minimizing the risk of transmitting animal spongiform encephalopathy via human or veterinary medicinal products, we check the CoO of raw material in respect to the specified animal source, the country of origin and the infectivity category. We neither store or process ruminant raw materials obtained from high infectivity tissues (IA) nor ruminant raw materials whose animal source originates from countries or regions with an undetermined risk (cat C/GBR IV).

	Safety Data
Toxic ingredients	• None
Basic composition	• See typical composition
Solvent content	• None
Safety data sheet required	• Not mandatorily required